

URGENT: DRUG RECALL

Estradiol Vaginal Inserts, USP, 10 mcg, 18's and 8's

(NDC: 68462-711-88 and 68462-711-71)

January 3, 2019

Good Afternoon,

This is to inform you of a voluntary Retail level recall involving the following:

NDC	Strength (10mcg)	Batch No	Expiration
68462-711-88	10	20180338	31.12.2019
68462-711-88	10	20180386	31.01.2020
68462-711-71	10	20180393	31.01.2020
68462-711-71	10	20180424	29.02.2020
68462-711-71	10	20180425	29.02.2020
68462-711-71	10	20180427	29.02.2020

Recall of these batches has been initiated for difficulty in pushing the plunger of the applicator.

Only the batch numbers listed are affected by this recall. This product was shipped to you between September 22, 2018 and October 26, 2018. Please be aware that there should be no ongoing disruption in supplying product due to this recall.

All recalled product will be processed by Qualanex. You will be receiving a packet from Qualanex with a recall letter, recall response form and detailed instructions regarding the return of any product related to this recall. If you do not receive this information by January 4, 2019 or have any questions please contact customerservice@glenmarkpharma.com.